



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

ILMO Products Company
7 Eastgate Drive, Jacksonville, IL 62650

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Chemical Testing
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

April 11, 2008

Issue Date:

December 8, 2024

Expiration Date:

February 28, 2027

Revision Date:

December 23, 2025


Accreditation No.:

61895

Certificate No.:

L24-940-R2

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjilabs.com*



Certificate of Accreditation: Supplement

ILMO Products Company

7 Eastgate Drive, Jacksonville, IL 62650
 Contact Name: Jerry Lanorio Phone: 217-243-7634

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Chemical	Compressed Gas Mixtures	Oxygen Concentration (%)	PurityPlus LWI 5.58	Paramagnetic	F1, F2	F
Chemical	Compressed Gas Mixtures	Trace Oxygen Concentration	PurityPlus LWI 5.60	Electrochemical	F1, F2	F
Chemical	Compressed Gas Mixtures	Total Hydrocarbon Concentration	PurityPlus LWI 5.7	Flame Ionization Detector	F1, F2	F
Chemical	Compressed Gas Mixtures	Trace Moisture Concentration	PurityPlus LWI 5.20	Electrolytic Cell	F1, F2	F
Chemical	Compressed Gas Mixtures	Trace Moisture Concentration	PurityPlus LWI 5.17	Cavity Ringdown Spectroscopy	F1, F2	F
Chemical	Compressed Gas Mixtures	Ethanol Concentration in Nitrogen	PurityPlus LWI 5.17 ILMO LWI-BAC	NDIR	F1, F2, F4	F
Chemical	Compressed Gas Mixtures	Trace Carbon Monoxide Concentration	PurityPlus LWI 5.17	NDIR	F1, F2	F
Chemical	Compressed Gas Mixtures	Carbon Dioxide Concentration (%)	PurityPlus LWI 5.17	NDIR	F1, F2	F
Chemical	Compressed Gas Mixtures	Gas Mixture Concentration	PurityPlus LWI 5.10	Gas Chromatography	F1, F3, F4	F
Chemical	Compressed Gas Mixtures	Ethanol Concentration in Nitrogen	PurityPlus LWI 5.17 ILMO LWI-FTIR	FT-IR	F1, F2, F4	F
Chemical	Compressed Gas Mixtures	Trace Gas Concentration	PurityPlus LWI 5.17 ILMO LWI-FTIR	FT-IR	F1, F2, F4	F
Chemical	Compressed Gas Mixtures	Trace Gas Concentration	PurityPlus LWI 5.16 ILMO GC-DID	GC-DID	F1, F2, F4	F

1. Location of activity:

Location

F

Conformity assessment activity is performed at the CABs fixed facility

Location

2. Flex Code:

- F0- Fixed scope item. No deviations allowed to the line item as identified, except for updating to the most recent version of an accredited standard method after verification.
- F1- Laboratory has the capability to test a new item, material, matrix, or product similar in composition to item, material, matrix, or product identified on the scope
- F2- Laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope
- F3- Laboratory has the capability to introduce a parameter/component/analyte to an accredited test method identified on the scope
- F4- Laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope
- F5- Laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using same technology or technique) identified on the scope